

**Test Report No. 7191237272-EEC20-03-WBH**  
dated 26 May 2020



PSB Singapore

**Add value.  
Inspire trust.**

**Note:** This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

**SUBJECT:**

Breathability test (Differential Pressure)

**CLIENT:**

MANN+HUMMEL FILTER TECHNOLOGY (S.E.A) Pte Ltd  
2 International Business Park  
#02-36, The Strategy  
609930 Singapore

Attn: Mr Terence Wong

**SAMPLE SUBMISSION DATE/ TEST DATE:**

11 May 2020/ 21 May 2020

**DESCRIPTION OF SAMPLES:**

One type of mask described as below was received:

Sample Description	Quantity
Community Face Mask	5 pieces

**METHOD OF TEST:**

BS EN 14683:2019 Medical face masks – Requirements and test methods  
-Clause 5.2.3 Breathability / Annex C Method for determination of breathability (differential pressure)

Airflow: 8L/min  
Tested area: 4.9 cm<sup>2</sup>  
Number of tested specimen: 5  
Number of test area per specimen: 5



Laboratory:  
TÜV SÜD PSB Pte. Ltd.  
No.1 Science Park Drive  
Singapore 118221

Phone : +65-6885 1333  
Fax : +65-6776 8670  
E-mail: enquiries@tuv-sud-psb.sg  
www.tuv-sud-psb.sg  
Co. Reg : 199002667R

Regional Head Office:  
TÜV SÜD Asia Pacific Pte. Ltd.  
1 Science Park Drive, #02-01  
Singapore 118221  
**TUV**<sup>®</sup>




**RESULTS:**

Sample Description: Community Face Mask

Test Specimen	Average Differential Pressure ( $\Delta P$ ) for 5 tested areas (Pa/cm <sup>2</sup> )	Requirement
1	17	The differential pressure of the medical face mask shall conform to the value given: Type I < 40 Pa/cm <sup>2</sup> Type II < 40 Pa/cm <sup>2</sup> Type IIR < 60 Pa/cm <sup>2</sup>
2	17	
3	17	
4	18	
5	18	

**REMARKS:**

1. The above test results relate to the samples as received.



William Wei  
Associate Engineer



Wong Bee Hui  
Product Manager  
Medical Health Services (NAM)


**APPENDIX:**



Photo: Community Face Mask

Test areas	Differential Pressure ( $\Delta P$ ) (Pa/cm <sup>2</sup> )				
	Specimen 1	Specimen 2	Specimen 3	Specimen 4	Specimen 5
1	16	16	16	17	17
2	17	16	17	17	16
3	16	16	16	16	15
4	20	20	19	20	21
5	18	19	19	19	20
Average	17	17	17	18	18

Table: Differential pressure for each tested area of the test specimen



Graph: Location of the areas of the mask for the differential pressure measurements

Please note that this Report is issued under the following terms :

1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
2. The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
3. Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
5. Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, No.1 Science Park Drive Singapore 118221.

July 2011

